

# **QUALITY ASSURANCE AUDIT REPORT**

## **North Texas Commission Ambient Air and Meteorological Monitoring**

**Prepared for:**

**North Texas Commission**

**8445 Freeport Parkway**

**Irving, TX 75063**

**Prepared by:**

**AECOM**

**13640 Briarwick Dr Suite 200**

**(78729)**

**P.O. Box 201088**

**Austin, TX 78720-1088**

**Conducted:**

**May 2023**

AECOM  
P.O. Box 201088  
Austin, TX 78720-1088  
13640 Briarwick Dr  
Suite 200  
Austin, TX 78729

## EXECUTIVE SUMMARY

On May 22<sup>nd</sup> through 24<sup>th</sup>, 2023, an audit team from the AECOM ambient air group in Austin, Texas conducted performance and technical system audits of the North Texas Commission (NTC) ambient air monitoring network. The audits provide an independent assessment of the monitoring program.

The monitoring program at NTC consists of continuous gas chromatographs (GC), volatile organic compound (VOC) canister collection systems, and meteorological sensors including wind speed, wind direction, and temperature.

The performance audit results indicate acceptable responses for measurement systems with the exceptions summarized below.

The temperature aspirator fan at Everman is not functioning properly. The temperature aspirator fan is spinning slowly resulting in reduced air flow over the temperature sensor. The fan should be replaced to ensure proper flow.

The temperature probe at Mansfield was outside of audit specification with an average error of -1.0° F.

Out of the 48 compounds being analyzed, nine compounds (ethylene, acetylene, isoprene, styrene, 1,3,5-trimethylbenzene, 1,2,4-trimethylbenzene, n-Decane, 1,2,3-trimethylbenzene, and n-undecane) were found to be outside of the audit objective of 70% - 130% recovery at several sites. In addition, the Benbrook, Decatur, Dish, Eagle Mountain Lake, and Mansfield sites had the following GC compound recoveries outside of the audit specification:

<b>Locations</b>	<b>Compounds</b>
Benbrook	Ethane Propylene
Decatur	Cis-2-Pentene
Dish	2-Methylhexane
Eagle Mountain Lake	Propylene
Mansfield	M&P-Xylene O-Xylene N-Nonane Isopropylbenzene n-Propylbenzene

These network GC audit results are comparable historically to other AECOM auto-GC audits. The GC audit results are contained in table ES-1. Technical systems audit results demonstrate satisfactory operational procedures for collecting valid data.

A performance evaluation (PE) sample is prepared by the AECOM QA group on a quarterly basis and submitted to the VOC laboratory for analysis. This performance evaluation sample contained known (spiked) concentrations of the target VOCs. A review of the sample recoveries for the spiked target VOCs shows that only one out of the forty-four compounds was not within the range of expected values (70-130%). Ethene had a recovery of 134.1%.

AECOM QA staff shared the performance evaluation results with the VOC laboratory, and no other corrective action was taken. We will continue to evaluate the compounds in our PE samples and work with the lab to resolve this discrepancy. GD Air's most recent performance evaluation canister results for the first quarter of 2023 are contained below in Table ES-2.

**Table ES-2. Audit Standard Results for all Network GCs**

Compound Name	CAS Number	Audit Conc (ppbc)	Benbrook		Decatur		Dish		Eagle Mountain Lake	
			Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery
Ethane	74-84-0	51.80	33.8	65.3%	41.6	80.3%	41.3	79.7%	40.8	78.8%
Ethylene	74-85-1	17.37	6.5	37.2%	11.6	66.8%	11.3	64.9%	10.2	58.7%
Propane	74-98-6	12.36	10.9	88.3%	10.3	83.3%	10.3	83.0%	9.4	76.2%
Propylene	115-07-1	13.09	9.0	68.7%	9.6	73.5%	9.6	73.7%	8.6	65.4%
Iso-Butane	75-28-5	16.97	18.2	107.5%	15.2	89.8%	15.9	93.7%	15.7	92.2%
N-Butane	106-97-8	17.13	19.1	111.3%	16.0	93.3%	16.5	96.1%	16.5	96.4%
Acetylene	74-86-2	8.73	5.3	60.7%	6.2	71.5%	6.0	69.1%	5.7	65.3%
Trans-2-Butene	624-64-6	16.65	18.2	109.4%	15.7	94.4%	16.2	97.4%	16.0	96.3%
1-Butene	106-98-9	16.65	18.1	108.5%	15.7	94.1%	16.2	97.5%	15.9	95.5%
Cis-2-Butene	590-18-1	17.62	19.2	109.1%	16.2	91.9%	17.0	96.3%	16.8	95.2%
Cyclopentane	287-92-3	21.21	23.0	108.3%	19.5	92.1%	20.1	94.6%	20.8	98.1%
Iso-Pentane	78-78-4	21.62	23.5	108.5%	20.1	93.2%	20.5	95.1%	20.8	96.2%
N-Pentane	109-66-0	21.62	23.9	110.4%	20.3	94.1%	20.8	96.1%	21.0	97.0%
1,3-Butadiene	106-99-0	16.65	16.4	98.4%	14.1	84.9%	16.3	97.7%	16.8	100.7%
Trans-2-Pentene	646-04-8	21.82	22.7	104.0%	19.3	88.7%	20.6	94.2%	21.1	96.8%
1-Pentene	109-67-1	21.82	20.1	92.2%	20.0	91.5%	20.4	93.3%	20.5	94.1%
Cis-2-Pentene	627-20-3	19.80	17.1	86.5%	13.6	68.7%	18.5	93.3%	18.8	94.8%
2,2-Dimethylbutane	75-83-2	25.45	26.8	105.3%	23.4	91.9%	23.5	92.3%	24.5	96.2%
2-Methylpentane	107-83-5	24.97	25.8	103.2%	21.9	87.5%	22.9	91.7%	23.8	95.2%
Isoprene	78-79-5	21.01	15.0	71.6%	14.1	67.3%	17.3	82.1%	17.7	84.1%
n-Hexane	110-54-3	26.18	26.2	99.9%	21.3	81.4%	21.8	83.3%	22.3	85.0%
Methylcyclopentane	108-87-2	25.70	24.2	94.2%	20.2	78.4%	19.0	74.0%	19.7	76.5%
2,4-Dimethylpentane	108-08-7	30.83	30.5	98.9%	27.4	88.9%	27.4	89.0%	30.0	97.2%
Benzene	71-43-2	25.94	23.1	88.9%	20.8	80.1%	21.6	83.1%	23.1	89.2%
Cyclohexane	110-82-7	26.18	27.3	104.5%	21.7	82.7%	22.9	87.3%	25.5	97.3%
2-Methylhexane	591-76-4	30.55	24.7	81.0%	22.4	73.2%	21.3	69.8%	22.1	72.5%
2,3-Dimethylpentane	565-59-3	29.70	32.4	109.1%	26.8	90.1%	27.5	92.6%	31.4	105.7%
3-Methylhexane	589-34-4	30.55	29.3	95.9%	25.0	81.9%	25.2	82.4%	27.8	91.0%
2,2,4-Trimethylpentane	540-84-1	35.23	31.5	89.4%	28.3	80.4%	27.0	76.8%	30.5	86.6%
n-Heptane	142-82-5	30.55	26.6	87.2%	25.3	82.9%	23.1	75.7%	27.0	88.5%
Methylcyclohexane	108-87-2	30.26	28.0	92.6%	25.0	82.7%	24.1	79.6%	26.0	86.0%
2,3,4-Trimethylpentane	565-75-3	33.94	31.4	92.4%	28.6	84.2%	27.9	82.1%	30.4	89.7%
Toluene	108-88-3	30.26	26.3	86.8%	24.5	81.1%	23.5	77.8%	26.2	86.5%
2-Methylheptane	592-27-8	34.26	29.90	87.3%	27.8	81.1%	26.6	77.5%	29.3	85.5%
3-Methylheptane	589-81-1	34.59	30.16	87.2%	28.2	81.6%	27.0	78.1%	29.4	85.1%
n-Octane	111-65-9	34.26	28.71	83.8%	28.7	83.7%	27.2	79.5%	29.4	85.8%
Ethylbenzene	100-41-4	34.26	26.40	77.1%	26.4	77.2%	26.3	76.8%	27.2	79.3%
M&P-Xylene	108-38-3	67.88	49.57	73.0%	50.9	75.0%	50.1	73.8%	52.8	77.7%
Styrene	100-42-5	32.00	19.21	60.0%	21.7	67.7%	22.9	71.7%	23.1	72.1%
O-Xylene	95-47-6	33.62	27.71	82.4%	26.8	79.6%	26.4	78.5%	28.6	84.9%
N-Nonane	111-84-2	37.09	30.50	82.2%	30.5	82.1%	29.8	80.2%	33.1	89.3%
Isopropylbenzene	98-82-8	36.36	37.37	102.8%	28.0	77.1%	27.7	76.2%	30.8	84.8%
n-Propylbenzene	103-65-1	34.91	25.60	73.3%	27.1	77.6%	26.5	75.9%	29.3	84.1%
1,3,5-Trimethylbenzene	108-67-8	37.82	25.77	68.1%	26.1	68.9%	26.7	70.7%	31.6	83.7%
1,2,4-Trimethylbenzene	95-63-6	38.55	25.58	66.4%	27.1	70.3%	28.3	73.3%	30.3	78.7%
n-Decane	124-18-5	42.83	26.96	62.9%	29.0	67.8%	29.2	68.2%	31.8	74.4%
1,2,3-Trimethylbenzene	526-73-8	39.64	22.20	56.0%	23.4	59.0%	24.9	62.7%	26.4	66.5%
n-Undecane	1120-21-4	46.22	25.42	55.0%	25.3	54.6%	27.1	58.6%	27.1	58.5%

<sup>a</sup> Compound order based on elution time.

**Table ES-1. (Continued) Audit Standard Results for all Network GCs**

Compound Name	CAS Number	Audit Conc (ppbc)	Elm Fork		Everman		Flower Mound		Godley	
			Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery
Ethane	74-84-0	51.80	40.9	79.0%	50.2	96.8%	41.7	80.5%	47.5	91.7%
Ethylene	74-85-1	17.37	12.4	71.5%	17.2	99.1%	12.3	70.9%	16.3	94.0%
Propane	74-98-6	12.36	10.3	83.3%	12.3	99.4%	10.5	85.2%	11.8	95.1%
Propylene	115-07-1	13.09	9.9	75.7%	13.2	101.0%	10.4	79.8%	11.6	88.4%
Iso-Butane	75-28-5	16.97	14.9	88.0%	17.9	105.3%	15.7	92.3%	16.6	97.7%
N-Butane	106-97-8	17.13	15.7	91.4%	18.3	106.7%	15.8	92.0%	17.1	99.7%
Acetylene	74-86-2	8.73	6.3	72.2%	8.0	91.8%	6.8	77.4%	7.9	90.9%
Trans-2-Butene	624-64-6	16.65	14.9	89.4%	17.9	107.4%	15.6	93.7%	16.6	99.5%
1-Butene	106-98-9	16.65	13.5	81.4%	17.8	106.7%	15.5	93.0%	16.4	98.3%
Cis-2-Butene	590-18-1	17.62	14.9	84.5%	18.7	105.9%	16.3	92.4%	17.1	97.3%
Cyclopentane	287-92-3	21.21	19.4	91.4%	22.2	104.5%	19.5	92.1%	20.5	96.7%
Iso-Pentane	78-78-4	21.62	19.8	91.4%	22.8	105.5%	20.0	92.4%	21.3	98.4%
N-Pentane	109-66-0	21.62	19.6	90.7%	22.7	105.2%	20.4	94.3%	21.2	98.3%
1,3-Butadiene	106-99-0	16.65	12.7	76.3%	17.2	103.6%	15.6	93.6%	16.1	96.8%
Trans-2-Pentene	646-04-8	21.82	19.8	90.5%	22.6	103.7%	20.2	92.4%	21.1	96.8%
1-Pentene	109-67-1	21.82	19.7	90.4%	22.4	102.6%	19.5	89.3%	20.4	93.7%
Cis-2-Pentene	627-20-3	19.80	16.3	82.5%	20.2	102.0%	17.8	89.9%	18.2	91.8%
2,2-Dimethylbutane	75-83-2	25.45	23.5	92.3%	26.3	103.4%	22.8	89.6%	24.6	96.6%
2-Methylpentane	107-83-5	24.97	22.8	91.2%	25.1	100.5%	23.1	92.6%	24.0	96.0%
Isoprene	78-79-5	21.01	14.6	69.6%	19.6	93.1%	16.5	78.6%	15.9	75.9%
n-Hexane	110-54-3	26.18	25.2	96.1%	26.4	100.8%	21.9	83.6%	25.2	96.4%
Methylcyclopentane	108-87-2	25.70	20.6	80.3%	25.4	98.7%	21.1	82.2%	25.1	97.8%
2,4-Dimethylpentane	108-08-7	30.83	25.9	83.9%	30.4	98.4%	26.3	85.2%	31.7	102.8%
Benzene	71-43-2	25.94	20.0	77.2%	26.3	101.6%	21.3	82.2%	23.2	89.5%
Cyclohexane	110-82-7	26.18	21.8	83.2%	28.2	107.8%	21.5	82.2%	25.6	97.8%
2-Methylhexane	591-76-4	30.55	22.6	74.1%	29.3	95.8%	23.7	77.7%	27.8	91.2%
2,3-Dimethylpentane	565-59-3	29.70	27.1	91.3%	32.7	110.2%	26.0	87.4%	30.5	102.6%
3-Methylhexane	589-34-4	30.55	25.2	82.5%	31.5	103.1%	25.6	83.7%	30.0	98.1%
2,2,4-Trimethylpentane	540-84-1	35.23	28.4	80.7%	34.8	98.7%	28.7	81.6%	33.7	95.6%
n-Heptane	142-82-5	30.55	25.1	82.3%	31.2	102.1%	25.3	82.8%	30.2	98.9%
Methylcyclohexane	108-87-2	30.26	25.1	83.0%	30.6	101.1%	25.2	83.3%	29.8	98.4%
2,3,4-Trimethylpentane	565-75-3	33.94	28.7	84.5%	34.4	101.3%	29.4	86.5%	33.8	99.6%
Toluene	108-88-3	30.26	23.7	78.3%	28.6	94.7%	24.6	81.1%	29.5	97.5%
2-Methylheptane	592-27-8	34.26	27.9	81.4%	34.3	100.0%	28.1	82.1%	33.3	97.2%
3-Methylheptane	589-81-1	34.59	27.8	80.5%	34.6	99.9%	28.5	82.5%	33.7	97.6%
n-Octane	111-65-9	34.26	28.1	82.0%	34.6	101.0%	28.0	81.9%	33.7	98.3%
Ethylbenzene	100-41-4	34.26	26.4	77.2%	33.0	96.5%	26.6	77.7%	32.0	93.5%
M&P-Xylene	108-38-3	67.88	50.6	74.6%	64.0	94.2%	50.5	74.4%	63.0	92.8%
Styrene	100-42-5	32.00	19.0	59.5%	27.3	85.2%	21.1	65.9%	28.0	87.4%
O-Xylene	95-47-6	33.62	25.0	74.3%	33.9	101.0%	26.3	78.3%	33.2	98.9%
N-Nonane	111-84-2	37.09	29.5	79.4%	39.0	105.2%	29.8	80.4%	37.3	100.6%
Isopropylbenzene	98-82-8	36.36	28.1	77.2%	37.2	102.2%	28.2	77.6%	36.8	101.3%
n-Propylbenzene	103-65-1	34.91	26.8	76.7%	35.0	100.4%	26.8	76.7%	35.2	100.9%
1,3,5-Trimethylbenzene	108-67-8	37.82	26.0	68.7%	35.1	92.8%	25.0	66.2%	38.4	101.5%
1,2,4-Trimethylbenzene	95-63-6	38.55	26.5	68.8%	31.8	82.6%	25.7	66.7%	38.0	98.6%
n-Decane	124-18-5	42.83	28.8	67.1%	37.3	87.0%	27.6	64.4%	38.9	90.9%
1,2,3-Trimethylbenzene	526-73-8	39.64	23.6	59.6%	30.2	76.2%	23.1	58.2%	35.6	89.8%
n-Undecane	1120-21-4	46.22	26.9	58.3%	31.2	67.4%	25.6	55.5%	40.2	87.0%

<sup>a</sup> Compound order based on elution time.

**Table ES-1. (Continued) Audit Standard Results for all Network GCs**

Compound Name	CAS Number	Audit Conc (ppbc)	Kennedale		Mansfield		Rhome		Rushing		UTA	
			Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery	Post Processed ppbc	Percent Recovery
Ethane	74-84-0	51.80	50.7	97.9%	38.7	74.8%	47.7	92.0%	46.8	90.3%	48.3	93.2%
Ethylene	74-85-1	17.37	11.5	66.3%	10.0	57.7%	15.5	89.0%	11.7	67.3%	14.7	84.9%
Propane	74-98-6	12.36	12.5	101.5%	10.0	81.2%	10.9	88.2%	11.2	90.9%	11.4	92.2%
Propylene	115-07-1	13.09	10.7	81.5%	9.5	72.4%	9.8	74.6%	9.6	73.2%	10.9	83.3%
Iso-Butane	75-28-5	16.97	17.9	105.7%	16.8	98.8%	16.2	95.7%	18.3	108.1%	16.7	98.4%
N-Butane	106-97-8	17.13	18.5	108.1%	17.3	101.0%	16.6	96.9%	19.1	111.8%	17.2	100.2%
Acetylene	74-86-2	8.73	8.2	94.1%	6.3	71.7%	7.5	85.4%	6.1	69.5%	7.5	86.2%
Trans-2-Butene	624-64-6	16.65	17.2	103.2%	16.6	99.6%	16.9	101.4%	18.4	110.7%	16.5	99.3%
1-Butene	106-98-9	16.65	18.2	109.1%	16.2	97.5%	16.4	98.5%	18.2	109.5%	16.6	99.8%
Cis-2-Butene	590-18-1	17.62	18.2	103.1%	17.0	96.5%	17.3	98.2%	18.8	106.8%	17.2	97.4%
Cyclopentane	287-92-3	21.21	22.3	105.1%	20.6	97.1%	20.7	97.7%	23.2	109.1%	20.9	98.7%
Iso-Pentane	78-78-4	21.62	23.0	106.3%	21.1	97.7%	21.8	100.9%	23.6	109.4%	21.6	99.9%
N-Pentane	109-66-0	21.62	22.9	106.0%	21.1	97.5%	21.7	100.4%	23.4	108.1%	21.5	99.5%
1,3-Butadiene	106-99-0	16.65	17.2	103.5%	16.0	96.2%	16.2	97.2%	17.3	103.7%	16.6	99.7%
Trans-2-Pentene	646-04-8	21.82	21.1	96.8%	20.6	94.5%	20.9	95.9%	21.6	99.0%	21.2	97.2%
1-Pentene	109-67-1	21.82	23.2	106.2%	20.3	92.9%	20.9	95.6%	21.0	96.1%	21.2	97.2%
Cis-2-Pentene	627-20-3	19.80	16.8	85.0%	18.3	92.3%	18.7	94.3%	14.9	75.2%	19.0	96.1%
2,2-Dimethylbutane	75-83-2	25.45	26.6	104.4%	24.4	96.0%	24.9	97.7%	26.2	102.9%	25.0	98.3%
2-Methylpentane	107-83-5	24.97	25.6	102.6%	24.2	97.0%	24.3	97.3%	25.8	103.1%	24.3	97.3%
Isoprene	78-79-5	21.01	17.0	81.1%	17.1	81.5%	18.4	87.7%	16.7	79.7%	18.2	86.5%
n-Hexane	110-54-3	26.18	24.9	95.2%	20.5	78.2%	26.2	99.9%	27.0	103.0%	25.7	98.0%
Methylcyclopentane	108-87-2	25.70	24.3	94.7%	20.3	79.0%	20.0	78.0%	24.4	95.0%	25.1	97.7%
2,4-Dimethylpentane	108-08-7	30.83	28.4	92.0%	24.6	79.9%	32.1	104.3%	29.0	94.2%	32.4	105.0%
Benzene	71-43-2	25.94	24.3	93.5%	20.3	78.3%	23.7	91.2%	24.3	93.8%	23.9	92.2%
Cyclohexane	110-82-7	26.18	25.8	98.7%	21.1	80.7%	25.2	96.1%	25.7	98.0%	26.5	101.1%
2-Methylhexane	591-76-4	30.55	27.8	91.0%	21.9	71.8%	24.3	79.4%	25.0	81.8%	27.6	90.2%
2,3-Dimethylpentane	565-59-3	29.70	29.7	99.9%	23.8	80.3%	32.3	108.6%	29.3	98.8%	31.3	105.2%
3-Methylhexane	589-34-4	30.55	29.1	95.3%	23.8	77.8%	29.3	96.0%	27.6	90.3%	30.2	98.9%
2,2,4-Trimethylpentane	540-84-1	35.23	32.7	92.9%	26.8	76.1%	32.1	91.2%	31.4	89.2%	34.0	96.5%
n-Heptane	142-82-5	30.55	29.3	95.9%	23.3	76.1%	28.3	92.5%	27.7	90.8%	30.2	98.9%
Methylcyclohexane	108-87-2	30.26	29.2	96.4%	23.8	78.5%	27.5	90.8%	28.2	93.1%	29.7	98.2%
2,3,4-Trimethylpentane	565-75-3	33.94	33.1	97.4%	27.0	79.6%	32.1	94.7%	31.1	91.6%	34.7	102.2%
Toluene	108-88-3	30.26	28.9	95.6%	22.3	73.6%	28.5	94.1%	26.7	88.4%	29.1	96.3%
2-Methylheptane	592-27-8	34.26	32.6	95.2%	25.6	74.6%	31.9	93.2%	30.0	87.5%	33.7	98.3%
3-Methylheptane	589-81-1	34.59	33.1	95.6%	26.0	75.1%	32.3	93.3%	29.8	86.2%	34.0	98.4%
n-Octane	111-65-9	34.26	33.7	98.5%	25.7	75.1%	32.2	93.9%	30.6	89.3%	34.6	100.9%
Ethylbenzene	100-41-4	34.26	30.6	89.4%	24.2	70.5%	29.1	84.8%	28.4	82.9%	31.9	93.1%
M&P-Xylene	108-38-3	67.88	60.0	88.4%	45.2	66.6%	57.2	84.3%	55.2	81.3%	62.6	92.3%
Styrene	100-42-5	32.00	26.4	82.5%	18.9	59.1%	26.4	82.4%	24.5	76.5%	27.1	84.7%
O-Xylene	95-47-6	33.62	31.3	93.2%	23.4	69.6%	30.4	90.5%	29.2	87.0%	33.0	98.0%
N-Nonane	111-84-2	37.09	36.2	97.6%	26.0	70.0%	36.2	97.5%	33.4	90.2%	39.2	105.8%
Isopropylbenzene	98-82-8	36.36	32.4	89.1%	25.0	68.9%	31.9	87.7%	31.3	86.0%	35.9	98.6%
n-Propylbenzene	103-65-1	34.91	32.0	91.6%	23.7	68.0%	31.5	90.1%	29.8	85.2%	34.7	99.4%
1,3,5-Trimethylbenzene	108-67-8	37.82	30.0	79.2%	22.8	60.3%	32.0	84.7%	29.1	77.0%	37.0	97.8%
1,2,4-Trimethylbenzene	95-63-6	38.55	31.8	82.5%	23.0	59.7%	34.4	89.3%	31.7	82.3%	35.4	91.9%
n-Decane	124-18-5	42.83	34.3	80.1%	23.8	55.6%	35.8	83.5%	32.6	76.2%	38.0	88.8%
1,2,3-Trimethylbenzene	526-73-8	39.64	27.5	69.5%	20.6	51.9%	28.5	72.0%	27.2	68.6%	31.9	80.4%
n-Undecane	1120-21-4	46.22	30.8	66.7%	23.8	51.6%	34.1	73.9%	31.2	67.4%	33.0	71.4%

<sup>a</sup> Compound order based on elution time.

**Table ES-2. Results of Performance Standard for Off-Site Analytical Lab**

Compound Name	CAS Number	Input Concentration	Lab Results	Percent Recovery
1,1,1-Trichloroethane	71-55-6	3.6	3.3	92.2%
1,1,2,2-Tetrachloroethane	79-34-5	3.5	3.5	98.9%
1,1,2-Trichloroethane	79-00-5	3.6	3.7	103.2%
1,1-Dichloroethane	75-34-3	3.5	3.2	89.4%
1,1-Dichloroethene	75-35-4	3.5	3.3	93.7%
1,2,4-Trimethylbenzene	95-63-6	3.3	2.6	79.7%
1,2-Dibromoethane	106-93-4	3.5	3.5	102.1%
1,2-Dichloroethane	107-06-2	3.6	3.4	95.6%
1,2-Dichloropropane	78-87-5	3.6	3.5	98.3%
1,3,5-Trimethylbenzene	108-67-8	3.4	2.7	79.6%
1,3-Butadiene	106-99-0	7.0	7.2	102.5%
1-Butene	106-98-9	3.5	3.1	90.0%
1-Hexene	592-41-6	3.2	3.3	102.4%
1-Pentene	109-67-1	3.4	3.3	97.3%
2,2,4-Trimethylpentane	540-84-1	3.5	3.6	103.1%
4-Ethyltoluene (p-Ethyltoluene)	622-96-8	3.3	2.7	80.7%
Benzene	71-43-2	3.6	3.5	96.7%
Bromomethane	74-83-9	3.4	3.0	89.4%
c-1,3-Dichloropropene	10061-01-5	3.2	3.6	114.6%
Carbon tetrachloride	56-23-5	3.5	3.2	91.7%
Chlorobenzene	108-90-7	3.5	3.5	100.0%
Chloroform	67-66-3	3.5	3.3	95.5%
Chloromethane (Methyl Chloride)	74-87-3	3.6	3.3	89.4%
Cyclohexane	110-82-7	3.5	3.5	98.8%
Dichlorodifluoromethane (Freon-12)	75-71-8	3.5	3.1	87.7%
Ethane	74-84-0	21.2	19.2	90.6%
Ethene	74-85-1	7.0	9.4	134.1%
Ethylbenzene	100-41-4	3.5	3.4	98.0%
Methylene Chloride (Dichloromethane)	75-09-2	3.5	3.3	94.9%
m-Xylene & p-Xylene	106-42-3+108-38-3	6.9	6.7	97.7%
n-Butane	106-97-8	3.5	3.4	96.3%
n-Heptane	142-82-5	3.4	3.5	101.9%
n-Hexane	110-54-3	10.5	10.2	97.0%
n-Pentane	109-66-0	3.4	3.4	99.9%
o-Xylene	95-47-6	3.5	3.4	96.9%
Propane	74-98-6	3.4	3.4	97.6%
Propylene	115-07-1	6.9	5.8	83.5%
Styrene	100-42-5	3.4	3.1	90.9%
t-1,3-Dichloropropene	10061-02-6	3.3	2.9	89.1%
Tetrachloroethene	127-18-4	3.6	3.2	87.5%
Toluene	108-88-3	3.6	3.5	97.6%
Trichloroethene	79-01-6	3.4	3.5	100.8%
Trichlorofluoromethane (Freon-11)	75-69-4	3.6	3.1	86.9%
Vinyl Chloride	75-01-4	3.6	3.2	90.6%